



Corporate Accreditation No 63
Accredited for compliance with ISO/IEC 17025 - Testing



Analytical Report 285934

Issue Date: 09/06/2023
Issued By : Sydney Water Laboratory Services

Delivery Address: Sydney Water Corporation
[Redacted]
West Ryde NSW 2114

Telephone: [Redacted]
Email: [Redacted]

Attention: [Redacted]
Customer: Department of Planning and Environment
Customer ID: ZOEH_1

Address: [Redacted]
Telephone: [Redacted]
Email: [Redacted]

CONTENTS

- 1. Sydney Water Approved Signatory
- 2. Sample Summary
- 3. Analytical results
- 4. Comments
- 5. Laboratory QC results

Sydney Water Approved Signatory

[Redacted], Phycology Supervisor	[Redacted], Organics Analyst	[Redacted], Organics Senior Analyst
----------------------------------	------------------------------	-------------------------------------

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

SAMPLE SUMMARY

<u>Client Sample ID</u>	<u>Sample Number</u>	<u>Sampling Procedure</u>	<u>Date Sampled</u>	<u>Date Received</u>	<u>Date Authorised</u>	<u>Description</u>
234094	L23043840	1	23/05/2023	25/05/2023	08/06/2023	B1 (ENVIRONMENTAL WATER)
234095	L23043841	1	23/05/2023	25/05/2023	08/06/2023	B2 (ENVIRONMENTAL WATER)
234096	L23043842	1	23/05/2023	25/05/2023	08/06/2023	E1 (ENVIRONMENTAL WATER)
234097	L23043843	1	23/05/2023	25/05/2023	08/06/2023	E2 (ENVIRONMENTAL WATER)
234098	L23043844	1	23/05/2023	25/05/2023	08/06/2023	E3 (ENVIRONMENTAL WATER)
234099	L23043845	1	23/05/2023	25/05/2023	08/06/2023	E4 (ENVIRONMENTAL WATER)
234100	L23043846	1	23/05/2023	25/05/2023	08/06/2023	B3 (ENVIRONMENTAL WATER)
234101	L23043847	1	23/05/2023	25/05/2023	08/06/2023	E5 (ENVIRONMENTAL WATER)
234110	L23043848	1	23/05/2023	25/05/2023	29/05/2023	B1 (ENVIRONMENTAL WATER)
234111	L23043849	1	23/05/2023	25/05/2023	29/05/2023	B2 (ENVIRONMENTAL WATER)
234112	L23043850	1	23/05/2023	25/05/2023	29/05/2023	E1 (ENVIRONMENTAL WATER)
234113	L23043851	1	23/05/2023	25/05/2023	29/05/2023	E2 (ENVIRONMENTAL WATER)
234114	L23043852	1	23/05/2023	25/05/2023	29/05/2023	E3 (ENVIRONMENTAL WATER)
234115	L23043853	1	23/05/2023	25/05/2023	29/05/2023	E4 (ENVIRONMENTAL WATER)
234116	L23043854	1	23/05/2023	25/05/2023	29/05/2023	B3 (ENVIRONMENTAL WATER)
234117	L23043855	1	23/05/2023	25/05/2023	29/05/2023	E5 (ENVIRONMENTAL WATER)

Sampling procedures

- 1 Samples analysed as received.
- 2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
- 3 Samples collected as per FS procedures SAWI 070.
- 4 Results reported as received from WNSW.

ANALYTICAL RESULTS

Client Sample ID	234094	234095	234096	234097	234098	234099	234100	234101
Sampled Date	23/05/2023 09:40:00 AM	23/05/2023 10:35:00 AM	23/05/2023 11:25:00 AM	23/05/2023 12:00:00 PM	23/05/2023 12:50:00 PM	23/05/2023 01:20:00 PM	23/05/2023 03:00:00 PM	23/05/2023 03:20:00 PM
Sample Number	L23043840	L23043841	L23043842	L23043843	L23043844	L23043845	L23043846	L23043847

ALGAL

MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes

Blue Green ASU	ASU/mL	2657	7103	4157	3069	5577	708.4	2311	3759
Blue Green Biovol	mm3/L	1	1.89	1.67	0.887	3	0.221	0.697	0.893
Potentially Toxic Blue Green	cells/mL	1940	2710	0	5250	0	0	2490	0
Potentially Toxic Blue Green ASU	ASU/mL	112.4	76.2	0.00	147.6	0.00	0.00	69.9	0.00
Potentially Toxic Blue Green Biovol	mm3/L	0.168	0.075	0.000	0.146	0.000	0.000	0.069	0.000
Total Blue Green	cells/mL	1320000	3701000	2150000	1546000	2878000	376000	1188000	1979000

MA91 : Individual Species Total Count, Total BioVol, Total ASU

Algae Source*	N/A	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL
Date of Performance	DD/MM/YY	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00	08/06/23 00:00

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID	234110	234111	234112	234113	234114	234115	234116	234117
Sampled Date	23/05/2023 09:40:00 AM	23/05/2023 10:35:00 AM	23/05/2023 11:15:00 AM	23/05/2023 12:00:00 PM	23/05/2023 12:50:00 PM	23/05/2023 01:20:00 PM	23/05/2023 03:00:00 PM	23/05/2023 03:30:00 PM
Sample Number	L23043848	L23043849	L23043850	L23043851	L23043852	L23043853	L23043854	L23043855

ORGANICS

TC0049DW : Algal Toxins

Anatoxin-a(extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Anatoxin-a(intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Anatoxin-a(total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Cylindrospermopsin (extra cellular)	ug/L	0.08	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Cylindrospermopsin (intra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Cylindrospermopsin(total)	ug/L	0.08	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID	234110	234111	234112	234113	234114	234115	234116	234117
Sampled Date	23/05/2023 09:40:00 AM	23/05/2023 10:35:00 AM	23/05/2023 11:15:00 AM	23/05/2023 12:00:00 PM	23/05/2023 12:50:00 PM	23/05/2023 01:20:00 PM	23/05/2023 03:00:00 PM	23/05/2023 03:30:00 PM
Sample Number	L23043848	L23043849	L23043850	L23043851	L23043852	L23043853	L23043854	L23043855

ORGANICS

TC0049DW : Algal Toxins(Continued)

Microcystin RR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Nodularin (extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Date of Performance	DD/MM/YY	29/05/23	29/05/23	29/05/23	29/05/23	29/05/23	29/05/23	29/05/23	29/05/23

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS

C1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID	234110	234111	234112	234113	234114	234115	234116	234117
Sampled Date	23/05/2023 09:40:00 AM	23/05/2023 10:35:00 AM	23/05/2023 11:15:00 AM	23/05/2023 12:00:00 PM	23/05/2023 12:50:00 PM	23/05/2023 01:20:00 PM	23/05/2023 03:00:00 PM	23/05/2023 03:30:00 PM
Sample Number	L23043848	L23043849	L23043850	L23043851	L23043852	L23043853	L23043854	L23043855

ORGANICS

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)

C2	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
dcGTX2	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX3	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
dcNeo	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcSTX	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX2	ug/L	<1	<1	<1	<1	<1	<1	<1	<1
GTX3	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX4	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX5	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX6	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
Neosaxitoxin	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3

* Indicates NATA accreditation does not cover the performance of this service

Client Sample ID		234110	234111	234112	234113	234114	234115	234116	234117
Sampled Date		23/05/2023 09:40:00 AM	23/05/2023 10:35:00 AM	23/05/2023 11:15:00 AM	23/05/2023 12:00:00 PM	23/05/2023 12:50:00 PM	23/05/2023 01:20:00 PM	23/05/2023 03:00:00 PM	23/05/2023 03:30:00 PM
Sample Number		L23043848	L23043849	L23043850	L23043851	L23043852	L23043853	L23043854	L23043855
ORGANICS									
TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)									
Saxitoxin	ug/L	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4
Date of Performance	DD/MM/YY	25/05/23	25/05/23	25/05/23	25/05/23	25/05/23	25/05/23	25/05/23	25/05/23

COMMENTS

<u>Sample ID</u>	<u>Comment Level</u>	<u>Method</u>	<u>Test</u>	<u>Comment</u>
L23043840	Method	MA91	-	Debris present in the sample.
L23043841	Method	MA91	-	Debris present in the sample.
L23043842	Method	MA91	-	Debris present in the sample.
L23043843	Method	MA91	-	Debris present in the sample.
L23043844	Method	MA91	-	Sample received unpreserved/ partially preserved, results may be compromised. Debris present.
L23043845	Method	MA91	-	Debris present in the sample.
L23043846	Method	MA91	-	Debris present in the sample.
L23043847	Method	MA91	-	Sample received unpreserved/ partially preserved, results may be compromised. Debris present.

* Indicates NATA accreditation does not cover the performance of this service

LABORATORY QC RESULTS

N/A - Not Applicable

PQL - Practical Quantitation Limit

LOQ - Limit of Quantification

RPD - Relative Percent Difference

SPIKE/Positive Control - Addition of a known amount and concentration

Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1

Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1

Duplicate Precision = Not calculated - Result is outside test range

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Anatoxin-a(extracellular)						
<0.1 ug/L	<0.1	100 50.0 - 120.0 ug/L	110 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Anatoxin-a(intracellular)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Anatoxin-a(total)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Cyindrospermopsin (extracellular)						
<0.05 ug/L	<0.05	89 50.0 - 120.0 ug/L	91 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Cyindrospermopsin (intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Cyindrospermopsin(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Microcystin LR(extracellular)						
<0.05 ug/L	<0.05	73 50.0 - 120.0 ug/L	70 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin LR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin LR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(extracellular)						
<0.05 ug/L	<0.05	85 50.0 - 120.0 ug/L	84 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin YR(extracellular)						
<0.05 ug/L	<0.05	86 50.0 - 120.0 ug/L	82 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Microcystin YR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin YR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Nodularin (extracellular)						
<0.1 ug/L	<0.1	77 50.0 - 120.0 ug/L	76 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Nodularin (intracellular)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Nodularin (total)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0061DW C1						
<0.5 ug/L	<0.5	110 70.0 - 130.0 ug/L	91 % Recovery 50.0 - 130.0 % Recovery	<0.5	<0.5	B 0.0 - 0.0 %
TC0061DW C2						
<0.3 ug/L	<0.3	110 70.0 - 130.0 ug/L	110 % Recovery 50.0 - 130.0 % Recovery	<0.3	<0.3	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0061DW dcGTX2						
<0.5 ug/L	<0.5	99 <i>70.0 - 130.0 ug/L</i>	75 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW dcGTX3						
<0.3 ug/L	<0.3	98 <i>70.0 - 130.0 ug/L</i>	88 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW dcNeo						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	51 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW dcSTX						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	58 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX1						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	84 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX2						
<1 ug/L	<1	110 <i>70.0 - 130.0 ug/L</i>	110 % Recovery <i>50.0 - 130.0 % Recovery</i>	<1	<1	B <i>0.0 - 0.0 %</i>
TC0061DW GTX3						
<0.5 ug/L	<0.5	100 <i>70.0 - 130.0 ug/L</i>	110 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0061DW GTX4						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	120 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW GTX5						
<0.5 ug/L	<0.5	80 <i>70.0 - 130.0 ug/L</i>	85 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX6						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	120 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW Neosaxitoxin						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	77 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW Saxitoxin						
<0.4 ug/L	<0.4	110 <i>70.0 - 130.0 ug/L</i>	74 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.4	<0.4	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

Extra Note:

F: Blank is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE Anatoxin-a(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcNeo B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcSTX B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX4 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX5 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX6 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Neosaxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Saxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

* Indicates NATA accreditation does not cover the performance of this service